

DEMULATOR, CHIP AND METHOD FOR DIGITALLY DEMODULATING AN FSK SIGNAL

CROSS-REFERENCE TO RELATED APPLICATION

This application claims the benefit of U.S. provisional application
5 Serial No. 60/538,178, filed January 22, 2004 and entitled "Frequency Shift Keying
Demodulation Methods for Wireless Biomedical Implants."

STATEMENT REGARDING FEDERALLY SPONSORED RESEARCH OR DEVELOPMENT

This invention was made with Government support from The
10 National Institute of Health under Contract No. NIH-NINDS-N01-NS-9-2304. The
Government has certain rights in the invention.

BACKGROUND OF THE INVENTION

1. Field of the Invention

This invention relates to demodulators, chips and methods for
15 digitally demodulating FSK signals.

2. Background Art

The following references may be referenced herein:

- [1] P.R. Troyk, I.E. Brown, W.H. Moore and G.E. Loeb,
"Development of BION Technology for Functional Electrical
20 Stimulation: Bidirectional Telemetry," 23rd IEEE-EMBS
Conference Proceedings, Vol. 2, pp. 1317-1320, 2001.
- [2] D.G. Galbraith, M. Soma, and R.L. White, "A Wideband
Efficient Inductive Transdermal Power and Data Link with

- Coupling Insensitive Gain," IEEE TRANS. BIOMED. ENG.
Vol. 34, pp. 265-275, April 1987.
- 5 [3] C.M. Zierhofer, E.S. Hochmair, "The Class-E Concept for
Efficient Wide-band Coupling-insensitive Transdermal Power
and Data Transfer," IEEE 14th EMBS Conference Proc.,
Vol. 2, pp. 382-383, 1992.
- [4] P.R. Troyk and M. Edington, "Inductive Links and Drivers
for Remotely-powered Telemetry Systems," Antennas and
Propagation Symposium, Vol. 1, pp. 60-62, 2000.
- 10 [5] C. Polk, E. Postow, "Handbook of Biological Effects of
Electromagnetic Fields," Chap. 2, CRC PRESS, 1986.
- [6] J.A. Von Arx and K. Najafi, "On-Chip Coils with Integrated
Cores for Remote Inductive Powering of Integrated
Microsystems," TRANSDUCERS 97, pp. 999-1002, June 1997.
- 15 [7] M. Ghovanloo, K. Beach, K.D. Wise, and K. Najafi, "A
BiCMOS Wireless Interface Chip for Micromachined
Stimulating Microprobes," IEEEEMBS Special Topic
Conference on Microtechnologies in Medicine and Biology
Proceedings, pp. 277-282, Madison-Wisconsin, May 2002.
- 20 [8] M. Ghovanloo and K. Najafi, "Fully Integrated Power
Supply Design for Wireless Biomedical Implants," IEEE-
EMBS Special Topic Conference on Microtechnologies in
Medicine and Biology proceedings, pp. 414-419, Madison-
Wisconsin, May 2002.
- 25 [9] M. Ghovanloo and K. Najafi, "A High Data Transfer Rate
Frequency Shift Keying Demodulator Chip for the Wireless
Biomedical Implants," IEEE 45th Midwest Symposium on
Circuits and Systems Proc., Tulsa-Oklahoma, August 2002.
- 30 [10] M. Ghovanloo and K. Najafi, "A Wideband Frequency-Shift
Keying Wireless Link for Inductively Powered Biomedical
Implants," IEEE TRANS. ON CIRCUITS AND SYSTEMS I, Vol.
51, No. 12, pp. 2374-2383, Dec. 2004.

- [11] M. Ghovanloo and K. Najafi, "A Modular 32-site Wireless Neural Stimulation Microsystem," IEEE JOURNAL ON SOLID-STATE CIRCUITS, Vol. 39, No. 12, pp. 2457-2466, Dec. 2004.

5 The inductive link between two magnetically-coupled coils is now one of the most common methods to wirelessly transfer power and data from the external world to implantable biomedical devices such as pacemakers and cochlear implants [1-3]. However, this is not the only application of data and power transfer via inductive coupling. Radio-frequency identification (RFID), remote sensing, and
10 MEMS are among a few other fields that can benefit highly from this method [4]. Achieving high power transfer efficiency, high data transfer bandwidth, and coupling insensitivity are some of the challenges that one would face in the design of such systems.

 Some of the biomedical implants, particularly those which interface
15 with the central nervous system, such as cochlear and visual prostheses, need large amounts of data to simultaneously interface with a large number of neurons through multiple channels. In a simplified visual implant for example, a minimum reasonable resolution of only 32×32 pixels for an image requires 10-bits for addressing, 8-bits for 256 gray levels, and 2-bits for polarity and parity-checking.
20 If one considers the human-eye natural bandwidth of 60 frames/sec, then 1.23 Mbps need to be transferred to this implant just as pure data. Therefore, a high data-rate receiver circuitry that can establish an efficient wireless link between the implant and the external units is highly needed.

 In broadband wireless communications such as IEEE 802.11a
25 standard for wireless LAN application, baud rates as high as 54 Mbps have been achieved at the expense of increasing the carrier frequency up to 5.8GHz, giving a data-rate to carrier-frequency ratio of only 0.93%. In other words, each data bit is carried by 107.4 carrier cycles. On the other hand, the maximum carrier frequency for biomedical implants is limited to a few tens of MHz due to the coupled coils'
30 self-resonant frequency, more power loss in the power transfer circuitry, and

excessive power dissipation in the tissue, which increases as square of the carrier frequency [5]. Therefore, a desirable goal is to transfer each data bit with a minimum number of carrier cycles to maximize the data-rate to carrier-frequency ratio and minimize the amount of power consumption.

5 So far, amplitude shift keying (ASK) data modulation has been commonly used in biomedical implants because of its fairly simple modulation and demodulation circuitry [1, 3, 4, 6, 7]. This method, however, faces major limitations for high-bandwidth data transfer, because high-bandwidth ASK needs high order filters with sharp cut-off frequencies, whose large capacitors cannot be
10 easily integrated in this low-frequency range of RF applications. A remedy that has been proposed in the so called suspended carrier modulation [1, 3, 4] boosts the modulation index up to 100% to achieve high data rates with low-order integrated filters at the expense of 50% reduction in the transferred power.

FSK data modulation technique has been was utilized for wirelessly
15 operating the University of Michigan micromachined stimulating 3D-microprobes, shown in Figures 1a and 1b, which are targeted at a 1024-site wireless stimulating microsystem for visual and auditory prostheses [7,8]. A receiver coil 10, hybrid components 11 and a telemetry interface chip 12 are enclosed within a hermetic package 13. STIM 2/2B probe shanks 14 extend from a substrate 15 which supports
20 electronic circuitry 16 thereon. This implantable microsystem consists of two major parts. First, a series of active (with circuitry) or passive (without circuitry) micromachined stimulating probes that are vertically mounted on a micromachined platform and second, a wireless interface chip that receives data and power through electromagnetic coupling and provides the entire system with regulated power,
25 synchronization pulses and a serial data bit-stream.

FSK Data Transfer

FSK is one of the most common modulation techniques for digital communication, which simply means sending binary data with two frequencies f_0 and f_1 , representing digital "0" and "1" respectively. The resultant modulated signal

can be regarded as the sum of two complementary 100% amplitude-modulated signals at different carrier frequencies as shown in Fig. 2a.

$$f(t) = f_0(t) \sin(2\pi f_0 t + \phi) + f_1(t) \sin(2\pi f_1 t + \phi) \quad (1)$$

In the frequency domain, the signal power is centered at two carrier frequencies, f_0 and f_1 , as shown in Fig. 2b. Since $f_0(t)$ and $f_1(t)$ can have the same amplitude, an excellent characteristic of the FSK modulation for wireless biomedical implants is that the transmitted power is always constant at its maximum level irrespective of f_0 and f_1 or the data content:

$$|f_0(t)| = |f_1(t)| = V_m \Rightarrow V_{rms}(f) = \frac{1}{\sqrt{2}} V_m \quad (2)$$

Another difference between the FSK and ASK is that in ASK data transmission the receiver tank circuit frequency response should have a very high quality factor (Q), centered at the carrier frequency to get enough amplitude variation for data detection. However, in FSK data transmission, the pass band should be centered between f_0 and f_1 with a low Q to pass enough power of both carrier frequencies. This is an advantage for the FSK technique because in the biomedical implant applications, the quality factor of the receiver coil is inherently low particularly when the implant receiver coil is integrated and its high resistivity is unavoidable [6]. The FSK signal is much less susceptible to the coupled coils misalignment and motion artifacts which are two major problems in biomedical implants that adversely affect the amplitude of the received signal.

Synchronization of the receiver with the transmitter is however easier in the ASK systems. Because the receiver internal clock signal can be directly derived by stepping down the constant transmitter carrier frequency [6, 7]. In FSK data transfer, the internal clock with constant frequency can be derived from a combination of the two carrier frequencies (f_0 and f_1) based on the data transfer protocol or synchronization patterns.

One wants to maximize the data-rate to carrier frequency ratio. Therefore, a particular protocol was devised for the FSK data transfer with the data-rate as high as f_I with f_0 twice as f_I . In this protocol, the digital bit "1" is transmitted by a single cycle of the carrier f_I and the digital bit "0" is transmitted by two cycles of the carrier f_0 as shown in Fig. 2a. The transmitter frequency switches at a small fraction of a cycle and only at zero crossings. This leads to a consistent data transfer rate of f_I Bits/sec. As a result, if one considers the average carrier frequency to be $(f_0 + f_I)/2$, then the data-rate to carrier frequency ratio can be as high as 67%. It is also useful to notice that any odd number of consecutive f_0 cycles in this protocol is an indication of data transfer error.

The following U.S. patents are related to the invention: 5,684,837; 4,616,187; 3,611,298; 3,623,075; 4,021,744; 3,979,685; 3,846,708; 3,908,169; 5,533,061; 4,115,738; 3,660,771; 5,550,505; 4,551,846; 3,600,680; 5,399,333; 4,488,120; 5,649,296; 4,485,347; 4,368,439; 4,825,452; 6,144,253; 4,103,244; 4,987,374; 5,748,036; 5,155,446; 4,568,882; 4,752,742; 5,245,632; 4,486,715; 4,533,874; 4,529,941; 6,359,942; 5,724,001; 6,038,268; 3,636,454; 6,501,807; 3,539,828; 4,010,323; 5,953,386; 4,773,085; 3,512,087; 3,501,704; 3,947,769; 5,053,717; 3,614,639; 3,949,313; 4,485,448; 5,436,590; 5,394,109; 4,716,376; 5,309,113; 3,427,614; 5,583,180; 4,363,002; 4,513,427; 6,307,413; 5,317,309; 6,122,329; 3,740,669; 4,451,792; 5,781,064; 5,329,258; 3,773,975; 3,991,389; 5,621,755; 6,366,135; 5,105,466; 3,372,234; and 4,543,953.

References [9] and [10] disclose a high-rate frequency shift keying (FSK) data transfer protocol and demodulator circuit for wirelessly operating biomedical implants in need of data transfer rates above 1 Mbit/Sec. The demodulator circuit receives the serial data bit stream from an FSK carrier signal in 2-20 MHz range, which is used to power the implant through inductive coupling.

The data detection technique used for the FSK demodulation is based on measuring the period of each received carrier cycle. If the period is higher than a certain value, a digital "1" bit is detected and otherwise a digital "0" is received. Time measurement is provided by charging a capacitor with a constant current

source and monitoring its voltage. Charging and discharging of this capacitor is synchronized with the FSK carrier signal. If the capacitor voltage is higher than a certain value, a digital "1" bit is detected and otherwise a digital "0" is received. This comparison can be done in two ways:

- 5 1. Fully differential FSK demodulator (FDFSK): Charging two unequal capacitors with different currents and compare their voltages with a hysteresis comparator. This is like comparing two capacitive timers with different time constants.
2. Referenced Differential FSK Demodulator (RDFSK): Generating a
10 reference voltage and comparing it with a charging capacitor voltage.

Such a demodulator circuit is usually a part the analog portion of a mixed signal chip and, in some cases, it is the only analog block on the chip. Analog blocks usually occupy more area than their digital counterparts. Specifically, the common FSK demodulation techniques need some kind of analog
15 filtering down the signal path, which consume even more chip area due to the above-noted low-end RF application. Analog circuits are more susceptible to process and temperature variations and their design becomes more challenging with the trend towards smaller feature size and lower power supply voltage. Therefore, a fully digital demodulator can save a lot of chip area, make the system more robust, and
20 ease many of the above problems.

SUMMARY OF THE INVENTION

An object of the present invention is to provide an improved demodulator, chip and method for digitally demodulating an FSK signal wherein at least one of the above-noted problems associated with the prior art is at least
25 partially solved.

In carrying out the above object and other objects of the present invention, in a high data rate receiver for an FSK data transmission system having a data transfer protocol, a digital demodulator for demodulating an FSK signal

having a data rate from a carrier signal having a pair of carrier frequencies is provided. The digital demodulator generates a serial data bit stream based on the FSK carrier signal and generates a synchronized constant frequency clock signal from the carrier frequencies based on the data transfer protocol for sampling the serial data bit stream. Both the data transfer protocol and the demodulator are fully digital to make the system robust.

The data rate may be greater than one million bits per second.

The data rate may approximate the carrier frequencies.

The carrier frequency may be less than about 25 megahertz and more than about 1 megahertz.

One of the carrier frequencies may be approximately twice the other carrier frequency so that a duration of each data bit is substantially the same, independent of its value.

The demodulator may also detect an error in the FSK carrier signal based on the protocol and provides a corresponding error signal.

The demodulator may also digitally measure the period of each received positive half cycle of the FSK carrier signal to obtain a series of pulses which distinguish between long and short FSK carrier cycles.

The demodulator may include an n-bit counter that runs with a clock time-base, f_{TB} , having a substantially constant frequency at a rate substantially higher than the FSK carrier frequencies, f_i and f_o , to digitally measure the periods.

The system may be a magnetically powered wireless system.

The receiver may be a wireless biomedical implant.

Further in carrying out the above object and other objects of the present invention, an FSK demodulator chip for an FSK data transmission system having a fully digital data transfer protocol is provided. The chip includes a substrate and a digital demodulator formed on the substrate for demodulating an FSK signal having a data rate from a carrier signal having a pair of carrier frequencies. The demodulator generates a serial data bit stream based on the FSK carrier signal and generates a synchronized constant frequency clock signal from the carrier frequencies based on the data transfer protocol for sampling the serial data bit stream. The demodulator is fully digital to minimize the amount of surface area occupied by the demodulator on the substrate.

The data rate may be greater than one million bits per second.

The data rate may approximate the carrier frequencies.

The carrier frequency may be less than about 25 megahertz and more than about 1 megahertz.

One of the carrier frequencies may be approximately twice the other carrier frequency so that a duration of each data bit is substantially the same, independent of its value.

The demodulator may also detect an error in the FSK carrier signal based on the protocol and provides a corresponding error signal.

The demodulator may also digitally measure the period of each received positive half cycle of the FSK carrier signal to obtain a series of pulses which distinguish between long and short FSK carrier cycles.

The demodulator may include an n-bit counter that runs with a clock time-base f_{TB} , having a substantially constant frequency at a rate substantially higher than the FSK carrier frequencies, f_1 and f_0 , to digitally measure the periods.

The system may be a magnetically powered wireless system.

Still further in carrying out the above object and other objects of the present invention, a method for demodulating an FSK signal having a data rate from a carrier signal having a pair of carrier frequencies in an FSK transmission system
5 having a digital data transfer protocol is provided. The method includes digitally measuring the period of each received positive half cycle of the FSK carrier signal to obtain a series of pulses which distinguish between long and short FSK carrier cycles. The method also includes digitally generating a serial data bit stream based on the FSK carrier signal and the series of pulses. The method further includes
10 digitally generating a synchronized constant frequency clock signal from the carrier frequencies based on the digital data transfer protocol and the series of pulses.

The above object and other objects, features, and advantages of the present invention are readily apparent from the following detailed description of the best mode for carrying out the invention when taken in connection with the
15 accompanying drawings.

BRIEF DESCRIPTION OF THE DRAWINGS

FIGURE 1a is a schematic perspective view of a wireless 3D-microstimulator array with hybrid-coil and interface chip mounted on a micromachined platform;

20 FIGURE 1b is a schematic perspective view of a hard-wired 3D-stimulator array [7,8];

FIGURES 2a and 2b are graphs which illustrate frequency shift keying in the time and frequency domains, respectively;

FIGURE 3a is a top plan view of a D-FSK demodulator prototype
25 chip which illustrates its floor plan;

FIGURE 3b is an enlarged portion of the chip of Figure 3a being labeled with circuit components;

FIGURE 4 is a generalized block diagram schematic view of a D-FSK demodulator constructed in accordance with one embodiment of the present invention;

FIGURE 5 is a circuit schematic view of a D-FSK demodulator constructed in accordance with one embodiment of the present invention;

FIGURE 6 illustrates a number of D-FSK demodulator simulated waveforms;

FIGURES 7a and 7b illustrate a number of D-FSK measured waveforms at 200 Kbps with f_0 and f_1 equal to 8 MHz and 4 MHz respectively; Figure 7a shows counter MSB C_2 , Data-out, CK_{in} , and Data-in [2 μ s/div]; Figure 7b shows Clock-out, Data-out, Carrier, and Data-in [500 ns/div];

FIGURE 8a is a schematic model of the combined series-parallel LC-tank with L_p used as the transmitter coil;

FIGURE 8b shows a simulated spectrum of the power amplifier output (V_{PA}) and the received signal (V_r); and

FIGURE 8c shows V_{PA} and V_r simulated waveforms in the time domain.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENTS

In general, an improved demodulator, chip and method for digitally demodulating an FSK signal is described herein. In this digital approach, the received FSK carrier is dealt with as a base-band signal. The data detection technique used here for the FSK demodulation is based on digitally measuring the

period of each received carrier cycle. If the period is higher than a certain value, a digital "1" bit is detected and otherwise a digital "0" is received.

As previously described, a simple method for time measurement in an analog circuit is charging a capacitor with a constant current source and monitoring its voltage [9, 10]. However, in a digital circuit one can count a constant frequency clock time-base (f_{TB}) at a rate several times higher than the carrier frequency. Most often this internal clock, which does not need to be synchronized with the carrier, is already available in the system for running other digital circuits or the processor. An n -bit counter runs while the carrier is "positive" and measures half of a carrier cycle. When the carrier goes "negative", the counter stops and a digital comparator decides whether a long or short carrier cycle has been received by comparing the count value with a constant reference. Then it resets the counter for measuring the duration of the next cycle.

The time-base period ($1/f_{TB}$) should be smaller than the time-difference between f_0 and f_1 half-cycles for the demodulator to be able to discriminate between these two frequencies. In other words f_{TB} should be chosen based on:

$$f_{TB} > \frac{2f_1f_0}{f_0 - f_1} \quad (3)$$

Therefore, the minimum width of the counter (n) should satisfy:

$$2^n > f_{TB} / f_0. \quad (4)$$

In order to make the demodulator circuit simpler and reduce dynamic power consumption, the digital comparator can be omitted by choosing f_0, f_1, f_{TB} and n so that:

$$f_0 > f_{TB} / 2^n > f_1. \quad (5)$$

In this case, the most significant bit (MSB) of the counter determines whether a long or short carrier cycle has been received and the constant reference value is equal to 2^n .

This algorithm was implemented in a prototype chip, which is shown in Figure 3 and fabricated in the AMI 1.5 μm 2M/2P standard-CMOS process. Even though processes with much smaller feature size are available, this process was chosen mainly because of its high voltage option, which is important for the stimulating blocks of the wireless stimulating microsystem shown in Figure 1a in order to overcome high site impedance when injecting current into the neural tissue. According to the FSK data-transfer protocol f_0 and f_1 were chosen equal to 8 MHz and 4 MHz, respectively, to achieve data-rates as high as 4 Mbps. By choosing these frequencies for the carrier, (3) sets a lower limit for the time-base clock ($f_{TB} > 16 \text{ MHz}$); and by choosing $n = 3$, (5) defines a new range for f_{TB} ($64 \text{ MHz} > f_{TB} > 32 \text{ MHz}$), which satisfies (3) as well. Therefore, a 7-stage ring-oscillator was designed to generate $f_{TB} = 49 \text{ MHz}$ at the middle of the above range. It should be noted that as long as f_{TB} is in the desired range, the phase noise and frequency variations of this oscillator do not affect the demodulation process.

Figure 4 shows a generalized block diagram schematic view of the D-FSK demodulator, Figure 5 shows a schematic diagram of the D-FSK demodulator, and Figure 6 shows sample simulation waveforms when a "00111100110011" data bit-stream is FSK modulated and applied to the D-FSK demodulator circuit {1}. The clock-regenerator circuit is a cross-coupled differential pair, which is directly connected to the receiver coil and turns the sinusoidal FSK carrier signal {2} into a similar square digital waveform called CK_{in} {3}. Utilization of a positive feedback in this circuit helps generate sharp output edges, which are important for precise timing of the carrier cycles. The internal ring-oscillator generates a time-base clock at f_{TB} , which runs the 3-bit counter after being gated by CK_{in} {4}. When CK_{in} is high, the counter is running, however it freezes when CK_{in} is low. The counter MSB (C_2) stays low during short (62.5 ns) carrier half-cycles when count < 4 , but goes high during long (125 ns) carrier half-

cycle when count $> 4 \{5\}$. The counter also resets when CK_{in} goes low during the 2nd carrier half-cycle to be ready for detecting the type of the next cycle.

The MSB of the counter (C_2) provides only a series of pulses, which discriminate between long and short FSK carrier cycles $\{5\}$. Therefore, it cannot be directly regarded as the received data bit stream. These pulses are fed into a digital block along with $CK_{in} \{3\}$ to generate the serial data output (Data-Out) $\{6\}$ and a constant frequency clock (Clock-Out) $\{7\}$. On every rising edge of the CK_{in} , a 2-bit shift register shifts in the C_2 pulses. Every 2 successive short cycles should be regarded as a "0" bit on Data-Out and every single long cycle indicates a "1" bit. Any odd number of short cycles is an indication of error according to the FSK protocol and activates the error flag. To generate a constant frequency clock, a T flip-flop indicates the number of successive zeros and another T flip-flop toggles on every long CK_{in} cycle or two successive short CK_{in} cycles. The resulting clock frequency is constant at $f_i/2$ irrespective of the data contents and the data values should be read at both rising and falling edges of the Clock-Out signal.

Measurement Results

The ring-oscillator, clock regenerator, and digital demodulator blocks were tested both individually and together as a digital FSK demodulator chip. With a 5V supply, the ring-oscillator generates f_{TB} at 50.5 MHZ, which is very close to the target value (49 MHZ) and inside the desired range. Figure 7 shows some of the measured waveforms, while f_0 and f_1 are set to 8 MHZ and 4 MHZ, respectively. All of the traces are labeled according to their names in the simulation and circuit diagrams. It can be seen that the FSK demodulator chip is functioning as expected from the simulations, up to 200 Kbps, which is the highest FSK modulation rate of our function generator (Agilent 33250A). A transmitter is capable of modulating up to 10 Mbps serial data bit-stream out of a fast digital I/O card (National Instruments DAQ-6534). Table 1 summarizes some of the specifications of the D-FSK demodulator chip.

TABLE 1
SPECIFICATIONS OF THE D-FSK DEMODULATOR CHIP

Process Technology	AMI 1.5 μm 2 M/2P Std-CMOS
Die size	2.2 mm x 2.2 mm
Circuit area	0.29 mm ² (2.41 mm x 0.12 mm)
Carrier frequency	4 ~ 20 MHZ
Maximum data rate	4 Mbps with 8 MHZ carrier
Time-base frequency	50.5 MHZ
Counter Width (n)	3-bits
Supply voltage	5V
Power dissipation	0.38 mW at 200 Kbps

The measured results that are reported in reference [10], page 2380 (also Fig. 12, page 2381) are at much higher rate than the ones mentioned here (*i.e.*, 2.5 MBit/sec vs. 200 kBit/sec). Also, reference [10], page 2381, Fig. 12, shows measured waveforms that are at much higher rates than the ones shown in Figs. 7a and 7b.

Wideband Inductive Link

Detailed design of the inductive link for biomedical implants is described in [2]-[3]. Almost all of these inductive links are designed for narrowband carrier signals, which is not the case in the proposed FSK modulation scheme. The main problem with a narrow bandwidth in high-speed FSK data transfer is inter-symbol interference, which is caused by the residual ringing that distorts the received carrier signal when the transmitter switches from one frequency to another. The Carlson's rule approximates the necessary bandwidth (*BW*) to include 98% of the total power of an FM signal:

$$BW \approx 2(\delta_{\max} + f_{i\max}) \quad (6)$$

where δ_{max} is the maximum frequency shift caused by modulation and f_{imax} is the maximum frequency content of the modulating signal. In the proposed FSK protocol, $\delta_{max} = f_i/2$ with respect to f_{avr} and the maximum data rate of f_i can be considered as a square waveform at $f_i/2$. Therefore, the main lobe of data spectrum
 5 has a maximum frequency of $f_{imax} = f_i$. By substituting these values in (6), a bandwidth of $BW \approx 3f_i$ is needed to include 98% of the FSK carrier power. However, it should be noted that contrary to the analog FM, here the goal is not a direct reconstruction of the data waveform, but correct detection of the data values. The experimental measurements show that an inductive link with half of the
 10 estimated bandwidth ($1.5 f_i$) can still provide an acceptable bit error rate (BER) in a system equipped with error detection circuitry.

The easiest way to increase the inductive link bandwidth is to lower the quality factor (Q) of the transmitter and receiver tank circuits by adding resistive components. However, the resulting increase in power dissipation especially in the
 15 implantable side is not desirable. Another method is to modify the transmitter output spectrum by adding two zeros at f_0 and f_1 , where the peaks of the FSK carrier spectral power are located. Figure 8a shows a simplified schematic diagram of the modified inductive link. The external parts of the system are replaced with an AC source and a source resistance (R_s), whereas the implant is replaced by a resistive
 20 load (R_L) and its parasitic input capacitance (C_{in}). A combination of both series and parallel LC-tank circuits are used to generate two zeroes across the FSK source output nodes at f_0 and f_1 . It can be shown analytically that this condition will be satisfied if $L_s C_s$ and $L_p C_p$ are chosen such that:

$$\frac{1}{2\pi\sqrt{L_p C_p}} = \frac{1}{2\pi\sqrt{L_s C_s}} = \sqrt{f_0 f_1} = f_m \quad (7)$$

25 and

$$\frac{1}{2\pi\sqrt{L_s C_p}} = f_1 \quad (8)$$

and

$$\frac{1}{2\pi\sqrt{L_P C_S}} = f_0 \quad (9)$$

In this series-parallel LC-tank combination, either L_P or L_S can be considered as the transmitter coil (L_t). The effect of the receiver coil (L_r) and its loading (C_r , C_{in} , and R_L) is neglected in the above calculations, because the mutual coupling between the two coils (M) is usually very small. Figure 8b shows the simulated spectrums of V_{FSK} and V_r , the voltages across the FSK source output and the receiver coil, when the circuit parameters are chosen based on Table 2. It can be seen that the zeros at f_0 and f_1 have provided a -3 dB bandwidth of about f_1 across L_r without any additional resistive components. The time-domain waveform of the received FSK carrier signal (V_r) in Figure 8c shows that the carrier frequency switches immediately with no residual ringing. Figure 8c also shows that f_0 and f_1 are suppressed at the FSK source output (V_{FSK}) and there are only spikes at the frequency switching points due to the other unsuppressed frequency components that become significant at these points.

TABLE 2
THE WIRELESS LINK CIRCUIT PARAMETERS

Parameter	Value	Comment
C_S	0.5 nF	Series tank capacitor
L_S	1 μ H	Series tank inductor
C_P	1 nF	Parallel tank capacitor
L_P	0.5 μ H	Parallel tank inductor
N_P	3	Parallel inductor turns
D_P	30 mm	Parallel inductor diameter
C_r	10 pF	Receiver tank capacitor
L_r^*	9 μ H	Receiver tank inductor
N_r	15	Receiver inductor turns
D_r	12 mm	Receiver inductor diameter
d_r	5 mm	Distance between L_r and L_P planar coils

Parameter	Value	Comment
M	100 nH	Mutual inductance between L_r and L_p
C_{in}	10 pF	Wireless chip parasitic input capacitance
R_L	1 k Ω	Wireless chip loading
R_S	50 Ω	Transmitter output resistance

5 * The receiver coil has a ferrite core.

Summary

A high-rate digital FSK data transfer protocol and demodulator circuit for wirelessly operating systems such as biomedical implants and radio frequency identification (RFID) tags has been described. However, it can be used for any other wireless application, which needs above 1 Mbps data transfer rate through an inductive link. The input clock (CK_{in}) is regenerated from the FSK carrier in 1 ~ 20 MHz range, which is used to power the implant as well. The serial data bit-stream and a constant frequency clock are then extracted from CK_{in} . The chip has been designed and fabricated in the AMI 1.5 μ m standard CMOS process, occupying an area of 0.29 mm². The prototype chip and its individual blocks have been simulated up to 2 Mbps and tested up to 200 Kpbs and were fully functional.

While embodiments of the invention have been illustrated and described, it is not intended that these embodiments illustrate and describe all possible forms of the invention. Rather, the words used in the specification are words of description rather than limitation, and it is understood that various changes may be made without departing from the spirit and scope of the invention.